

MPW-260

MPW-260R

MPW-260RH



MPW - 260

SPEED	6300	0
RF	2486	0
TIME	00:02:00	00:02:00

MPW MED. INSTRUMENTS

SHORT START STOP COVER BACK

PDF



FEATURES

effective ventilation system [MPW-260]

with cooling [MPW-260R]

with cooling and heating [MPW-260RH]

temperature +4°C ensured for maximum speed of each rotor [MPW-260R; MPW-260RH]

delayed start function

start when preselected temperature is reached [MPW-260R; MPW-260RH]

initial cooling without/with centrifuging, cooling after centrifuging, precise temperature stabilization [MPW-260R; MPW-260RH]

maintenance-free induction motor

modern software system

large graphic LCD [MPW-260RH] - blanking the screen after a period of inactivity

preselected and real values of speed, RCF, time shown on the display at the same time [MPW-260]

preselected and real values of speed, RCF, time and temperature shown on the display at the same time [MPW-260R; MPW-260RH]

100 user programs

10 acceleration/deceleration curves

programming of multi-section acceleration/deceleration curves

RPM/RCF adjustment

two modes of time calculation: from start key pressing or from reaching preselected speed

continuous operation mode - HOLD

SHORT time operation mode

manual centrifugal radius adjustment [with automatic RCF correction]

possibility of changing of parameters during centrifuging

automatic rotor identification

automatic lid opening

centrifuging parameters recording - MPW Editor - PC computer [USB]

menu language setting [PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ]

selected functions blocking, password protection

manual adjustment of density of samples >1,2g/cm³ [with automatic maximum speed correction]

stainless steel rotor chamber

SAFETY












unbalance sensor










lid locking during rotor running

start blocking at opened lid

emergency lid lock release

product conforming with the EN-61010-1 and EN-61010-2-020, EN-61010-2-101 standards

	S-LINE				M-LINE				
									
	MPW-54	MPW-55	MPW-56	MPW-150R	M-BASIC	M-UNIVERSAL	M-DIAGNOSTIC	M-SCIENCE	MPW-260
	120 ml	48 ml	120 ml	90 ml	120 ml / PRP	500 ml	500 ml	100 ml	500 ml
RCF max.	1 137, 3 122 x g	15 279 x g	3 341 x g	21 382 x g	2 469 x g	24 270 x g	4 830 x g	24 270 x g	24 270 x g
RPM max.	3 500, 5 800	14 500	6 000	15 000	4 000	18 000	6 000	18 000	18 000
	-	-	-	-20°C ÷ +40°C	-	-	-	-	-
Number of rotors (angle swing-out)	3/0	3/1	4/0	7/1	1	19/4	8/5	8/1	19/3
Rotor auto-recognition	-	-	-	•	-	•	•	•	•
Language	EN	EN	EN	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	EN	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ
Cover tightening	-	-	-	-	-	-	-	-	-
User programs	-	9	-	100	5	100	100	100	100

	M-LINE	L-LINE				XL-LINE		
								
	MPW-260R	MPW-260RH	MPW-351e	MPW-352	MPW-352R	MPW-352RH	MPW-380	MPW-380R
	500 ml	500 ml	800 ml	1 000 ml	1 000 ml	1 000 ml	3 000 ml	3 000 ml
RCF max.	24 270 x g	24 270 x g	3 509 x g	29 703 x g	30 065 x g	30 065 x g	31 876 x g	31 876 x g
RPM max.	18 000	18 000	4 500	18 000	18 000	18 000	18 000	18 000
	-20°C ÷ +40°C	-20°C ÷ +55°C	-	-	-20°C ÷ +40°C	-20°C ÷ +55°C	-	-20°C ÷ +40°C
Number of rotors (angle swing-out)	19/3	19/3	4/3	27/7	27/7	27/7	13/4	13/4
Rotor auto-recognition	•	•	-	•	•	•	•	•
Language	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	EN	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ	PL, EN, ES, IT, PT, DE, RU, SE, FR, CZ
Cover tightening	-	-	-	•	•	•	•	•
User programs	100	100	-	100	100	100	100	100



46 Boremlowska Street
04-347 Warsaw Poland

MPW MED. INSTRUMENTS
www.mpw.pl

E-mail: mpw@mpw.pl
Phone.: +48 22 610 56 21
Fax: +48 22 610 55 36

